

FLUKE

....

CE

DECLARATION OF CONFORMITY

Apparatus Identification

Fluke Networks DTX Series Cable Analyzers Models 1800, 1200 and LT

Apparatus Classification

Test and Measurement Equipment

Statement of Conformity

Based on sample product test results using appropriate standards, and in accordance with the following EC Directives Fluke Networks hereby declares the **DTX Series Cable Analyzers** to be in conformity with:

**EC Directive 89/336/EEC, Electromagnetic Compatibility and
EC Directive 73/23/EEC, Low Voltage (Amended 93/68/EEC).**

Sample Product Testing for EMC

Tested By Fluke Corporation
P.O. Box 9090
Everett, WA 98206-9090

Standards Used **EN 61326-1 :1997 +A1: 1998 +A2:2000** Electrical equipment for measurement, control and laboratory use.

Report ID Electromagnetic Compatibility Test Report for the **Fluke Networks DTX 1800 Cable Analyzer.**

Sample Product Testing for Safety

Tested by Fluke Corporation
PO Box 9090
Everett, WA 98206-9090

Standards used **EN 61010-1 : 2000 2nd edition**


Report ID Product Safety Conformance Verification Report for the **Fluke Networks DTX Series Cable Analyzers.**

Manufacturer

FLUKE Corporation
6920 Seaway Boulevard
PO Box 9090
Everett, WA 98206-9090
United States of America

EU Representative

FLUKE Europe B.V.
Science Park Eindhoven 5110
5692 EC SON Eindhoven
The Netherlands


Chris Odell, President
Fluke Networks


Jan Roetman, General Manager
Fluke Europe B.V.

DC20043-217

15-October-2004